

EVALUATION OF VIDAS® CLOSTRIDIUM DIFFICILE TOXIN A&B ON COLONIES



C. Eckert^{1,3}, P. Schaaf¹, C. Ménard², C. Elias², B. Burghoffer^{1,3}, V. Lalande^{2,3}, F. Barbut^{1,2,3}

¹National Reference Laboratory for *C. difficile*, UPMC Univ Paris 06, Paris, France

²Microbiology, Saint Antoine Hospital, AP-HP, Paris, France

³UPMC Univ Paris 06, GRC n° 2, Epidiff, Paris, France



Objective

Toxigenic culture (TC) is considered as a reference method for the diagnosis of *Clostridium difficile* infections (CDI). This method consists in isolating *C. difficile* on selective media and then assessing the *in vitro* toxins production from colonies. However determination of the toxigenic status of a strain using immunoenzymatic assays (IEA) on colonies is not always validated. The objective of this study was to assess the performance of VIDAS *Clostridium difficile* Toxin A&B (bioMérieux) (VIDAS) directly on colonies from chromID® *C. difficile* medium (bioMérieux) (chromID).

Materials and Methods

During a study where different selective media for *C. difficile* were compared, 36 strains were isolated on both chromID® *C. difficile* agar incubated anaerobically for 24 hours (chromID) and TCCA incubated anaerobically for 48 hours.

The *in vitro* determination of the cytotoxicity of isolates was performed by inoculating 2 to 5 colonies isolated on TCCA into BHI that was incubated for 5 days under anaerobic conditions. The supernatant from this culture was filtered and inoculated on MRC-5 cells. This method is referred as toxigenic culture (TC) (Figure 1A, 1B).

In parallel, the VIDAS *Clostridium difficile* Toxin A&B (VIDAS CDAB) test was performed on colonies from chromID (from primary culture or subculture to obtain sufficient colonies) (Figure 2A). Colonies were homogenized with water to reach a 3 McFarland density. VIDAS tests were then performed according to the manufacturer's instructions (Figure 2B). The VIDAS tests were repeated whenever results were equivocal ($VT \geq 0.13$ and < 0.37) or discordant (positive-TC and negative-VIDAS result or negative-TC and positive-VIDAS result).




Fig.1. A. MRC5 cells. B. cytopathic effect on MRC5 cells due to the toxin B




Fig 2 A. *C. difficile* colonies on chromID *C. difficile* agar. B. miniVIDAS (bioMérieux)

Results

Among the 36 strains studied, 27 (75%) were positive and 9 (25%) negative by TC.

There was sufficient colonies on chromID agar in 29 cases (80.6%) allowing to perform VIDAS assay on colonies from primary culture. In 7 cases (19.4%) a subculture was necessary to have sufficient colonies to reach a 3 McFarland density.

Results were equivocal in 6 cases (16.7%). After repeating the assay results remained equivocal in one case (2.8%). The sensitivity and specificity for VIDAS on colonies from chromID compared to TC were 70.4% [49.7-85.5] and 87.5% [46.7-99.3], respectively (Table 1).

Table 1. Performances of the VIDAS CDAB for the detection of toxigenic strains of *C. difficile*

	TC				VIDAS on colonies/TC
		+	-	N=	
VIDAS on colonies	+	19	1	20	Sensitivity (%) (CI 95%) 70.4 (49.7-85.5)
	-	8*	7	15	
	N=	27	8	35	Specificity (%) (CI 95%) 87.5 (46.7-99.3)
Equivocal result			1		

*7 Vidas assays were performed on primary culture

After resolving discordant results, the sensitivity and specificity for VIDAS on colonies from chromID compared to TC were 83.3% [61.8-94.5] and 100% [59.8-100], respectively (Table 2). Results were equivocal in 4 cases (11.1%).

Table 2. Performances of the VIDAS CDAB for the detection of toxigenic strains of *C. difficile* after resolving discrepant results

	TC				VIDAS on colonies/TC
		+	-	N=	
VIDAS on colonies	+	20	0	20	Sensitivity (%) (CI 95%) 83.3 (61.8-94.5)
	-	4	8	12	
	N=	24	8	32	Specificity (%) (CI 95%) 100 (59.8-100)
Equivocal result		3	1		

Conclusions

VIDAS *Clostridium difficile* Toxin A&B performed directly on colonies from chromID® *C. difficile* agar (24h) is a simple and rapid test to confirm the toxigenicity of the microorganism, when sufficient colonies are present on the medium. However, negative results must be confirmed by a reference test.